



Specification Sheet

Product: Colour coded buffer solution, pH4,
red

Code: J/2825/15

Grade: Buffer Solution

Application:

Expiry Phrase: Use within 2 yrs of opening

Test
pH @ 20°C

Guaranteed
≥ 3.98 and ≤ 4.02

Units
pH

Additional Information: The pH value is determined from the response of a calibrated glass electrode. The pH scale is that defined in BS1647:1984. The glass electrode is calibrated by measuring its response when immersed in reference buffer solutions. Reference buffer solutions are prepared according to the formulations in the British Standard and are traceable to NIST reference material

Mrs J M Wilson.
BSc C.Chem MRSC.
QA Laboratory Manager.

Friday 08th of April 2011 11:15:29 AM
Any queries please phone: +44(0)1509 555158.

