



## Specification Sheet

Product: Aqualine complete 5

Code: K/2000/15

Grade:

Application: Karl Fischer volumetric reagent

Expiry Phrase: Expires 2 yrs from packing date, see label.

**Test**  
Factor (mg/ml)

**Guaranteed**  
 $\geq 5.3$  and  $\leq 5.6$

**Units**  
mg/ml

Additional Information: A single component (composite) with a water equivalence of 5 mg water per ml. The factor result and specification apply at the time of testing only. The factor of this product will decrease over time

Mrs J M Wilson.  
BSc C.Chem MRSC.  
QA Laboratory Manager.

Wednesday 26th of January 2011 08:02:04 AM  
Any queries please phone: +44(0)1509 555158.

**ThermoFisher**  
SCIENTIFIC  
*The world leader in serving science*